



# SIVA INSTRUMENTS



QUALITY ASSURED COMPANY

ISO 9001 : 2000

## BREAKDOWN VOLTAGE TESTER

Model : HVO-50 50KV AC 50mA



### Technical Specifications

**Test Voltage:** 0-50KV AC.

**Accuracy** :  $\pm 2\%$  of reading  $\pm 2$  digits.

**Trip Current** : 50 mA Pre-Settable.

**Timer** : 0-999Seconds

**Auto/Manual Modes** : In Auto Mode Auto Voltage Setting, Auto Trip Settings & Timer Settings will have control over test. Push Buttons & Knobs are provided for settings.

**Dimmerstat** : Motorized.

**Display** : 3 1/2 DPM display for Voltage & Current  
3 digit for Timer. Same displays used for Preset Value settings.

**Controls** : Power ON, HV ON, HV OFF, Reset, Voltage Increase/Decrease, Audio Visual Failure & HV indicator.

**Input** : 230 VAC , 50Hz , single Phase  
(110V, 60Hz on request)

**Protections** : Zero Start Protection  
: HV Cover interlock.  
Open Ground Protection.

### Special Features

- Metal Cabinet.
- Selectable Trip Current.
- Epoxy Molded Transformer.
- Provided with safety features.
- Portable.
- Easy to Operate.

To keep pace with the latest technology Sivananda Electronics reserves the right to update it's system. Hence, information in the brochure is likely to change without notice.



## SIVANANDA ELECTRONICS

DEEPAK MAHAL, LAM ROAD, DEVLALI-422 401, NASHIK. MAHARASHTRA, INDIA.

Phone : 0253-2491423, 2491504, 2491816 E-mail : mkb@sivanandaelectronics.com Fax: 0253-2492291

MUMBAI : 208, MIDAS Sahar Plaza Complex, Andheri Kurla Road, Andheri (E) Mumbai 400059.

Ph: 28271046, 66430590, Fax : 28389232.

NEW DELHI : G -145, KALKAJI, NEW DELHI -110 019. ☐PHONE :26221582, 26434442

BANGALORE : 44/2, DICKENSON ROAD, BANGALORE - 560 042 ☐PHONE : 25585673

HYDERABAD : 102, ART PLAZA, YWCA LANE, NEAR HIRA HALL, BOGGOUULKUNTA, ABIDS, HYDERABAD-500001.

☐PHONE :040-30629539, e-mail:sivahyd@sify.com

MFRS. OF : ELECTRONIC INSTRUMENTS, METAL DETECTORS, TURNSTILES, ACCESS CONTROL