




SIVA INSTRUMENTS

Million Megohm Meter

Model :LS-3BD/ LS-3CD



An ISO 9001:2008 COMPANY



Specifications

- Test Voltage : - For LS-3BD = 63V / 100V / 500V DC settable
For LS-3CD = 100V / 250V / 500V / 1000V DC settable
 - Accuracy : $\pm 1\%$ of set Value ± 1 digit
- Insulation Resistance : - Accuracy - $\pm 2\%$
 - Range IR = 10Meg to 10e9 Megohm
- Separate Indication for Charge / Measure and Discharge mode

SIVANANDA ELECTRONICS

DEEPAK MAHAL, LAM ROAD, DEVLALI-422 401, NASHIK. MAHARASHTRA, INDIA. Phone : 0253-2491423, 2491504, 2491816 E-mail : mkb@sivanandaelectronics.com Fax: 0253-2492291

MUMBAI : 208, MIDAS Sahar Plaza Complex, Andheri Kurla Road, Andheri (E) Mumbai 400059.
Ph: 56430590, Fax : 28389232.

NEW DELHI : G -145, KALKAJI, NEW DELHI -110 019. ☎PHONE : 26221582, 26434442, FAX: 011-26434440

BANGALORE : 44/2, DICKENSON ROAD, BANGALORE - 560 042 ☎PHONE : 5585673

MFRS. OF : ELECTRONIC INSTRUMENTS, METAL DETECTORS, TURNSTILES, COMPUTERISED ATTENDANCE RECORDING & ACCESS CONTROL